

Department of Forensic Science  
Breath Alcohol Section  
Instrument Location History  
1/1/1990 To 8/26/2014

AMS  
pd

Instrument Serial Number: 010603

As of 26-Aug-14

Start Date	End Date	Agency Name
26-Aug-14		DFS Central Lab
31-Mar-14	26-Aug-14	Fairfax County PD
25-Sep-13	31-Mar-14	DFS Central Lab
14-Mar-13	25-Sep-13	Fairfax County PD
05-Nov-12	14-Mar-13	DFS Central Lab
24-May-12	05-Nov-12	Virginia Beach PD
27-Mar-12	24-May-12	DFS Central Lab
22-Nov-11	27-Mar-12	Richmond PD
13-Sep-10	22-Nov-11	DFS Central Lab
04-Jun-10	13-Sep-10	Intoximeters, Inc.
12-May-09	04-Jun-10	DFS Central Lab
18-Mar-09	12-May-09	Norfolk PD
04-Aug-08	18-Mar-09	DFS Central Lab

HHS gnd

# INTOX EC/IR II

## Quality Assurance Worksheet

Instrument Serial Number **010603** Worksheet Start Date **8/26/2014**

Location **Fairfax County Adult Detention Center**

Address **10520 Judicial Dr., Fairfax, VA 22030**

DFS Technician **Heather Stanton** License No. **18910**

☐ Laboratory ☒ On-Site

Site Specification: No detrimental environmental conditions exist. ☒

Instrument Barometer (mm HG) **747** Reference Barometer (mm HG) **755**  
**WARNING difference > 7**

Reference Barometer(RB)Serial # **009113** RB Calibration Due **4/28/2015**

### Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.293
0.300	0.295	0.286	0.303	Sample 2	0.292
Precision		sample min	sample max	Sample 3	0.291
0.002		0.291	0.293		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.098
0.100	0.098	0.095	0.101	Sample 2	0.098
Precision		sample min	sample max	Sample 3	0.098
0		0.098	0.098		

Dry gas standard Lot No. (with tank no.) **AG230501-37**

☐ Replaced dry gas standard (+O-ring)

☐ Installed at Location

☒ Removed to DFS-Central

Supplies  
Mouthpieces  
Certificates of Analysis  
Operator Worksheet  
Other:



Notes:

Instrument Serial Number

010603

Certification Date

☐ Calibrated☐ Certified☒ Measurement Assurance Check☒ Instrument Test☐ Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	


Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section


DFS Technician



Date

8/26/14

Issuing Analyst



Date

8/27/14



COMMONWEALTH OF VIRGINIA  
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS  
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED  INSTRUMENT, TEST,	NAME OF COURT  DFS
--	--------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY  STANTON, HEATHER, M		AGENCY  DFS Central Lab
DFS LICENSE NUMBER  18910	LICENSE EXPIRES  10/01/2014	DATE TEST CONDUCTED  08/26/2014
TEST EQUIPMENT NUMBER  010603		

RESULTS: TIME SAMPLE TAKEN 10:29 EDT  
SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS \_\_\_\_\_ DAY OF \_\_\_\_\_, 20 \_\_\_\_\_.

BREATH TEST OPERATOR

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS \_\_\_\_\_ SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS \_\_\_\_\_ OPERATOR'S SIGNATURE



# IntoxNet MIS Report

Report Generated 26 Aug 2014 at 11:02

*Handwritten initials/signature*

## Test Results

**Instrument Serial Number 010603**

**Test # 001376**      **Subject Test**

**Test Location 1** Department of  
Test Date 26 Aug 2014

**Test Location 2** Forensic Science  
Test Time 10:22  
Remote/Local Local

**Test Location 3**  
System Check Passed

**Operator's Last Name** STANTON  
**Agency** DFS Central Lab

**Operator's First Name** HEATHER

**Operator's Middle Initial** M

**Card Serial Number** 118910

**License Number** 18910

**Effective Date** 10/01/2012

**Expiration Date** 10/01/2014

**Subject's Last Name** INSTRUMENT

**Subject's First Name** TEST

**Subject's Middle Initial**

**Subject's Date of Birth** 00/00/0000

**Subject's Sex** Male

**Driver's License Number**

**Driver's License Expiration** 00/00/0000

**Driver's License State**

**Court Name** DFS

**End Date** 26 Aug 2014      **End Time** 10:30

**Result Time** 10:29

**Result Date** 26 Aug 2014

**Result** 0.00

**Data Type** DIAG

**Sample Value** Pass

**Sample Time** 10:23

**Data Type** BLK

**Sample Value** 0.000

**Sample Time** 10:23

**Data Type** CHK

**Sample Value** 0.098

**Sample Time** 10:24

**Data Type** BLK

**Sample Value** 0.000

**Sample Time** 10:26

**Data Type** SUBJ

**Sample Value** 0.000

**Sample Time** 10:26

**Data Type** BLK

**Sample Value** 0.000

**Sample Time** 10:27

**Data Type** SUBJ

**Sample Value** 0.000

**Sample Time** 10:29

**Data Type** BLK

**Sample Value** 0.000

**Sample Time** 10:30

**Standard Type** Dry Gas Std

**Standard Value** 0.098

**Standard Lot Number** AG230501-37

**Standard Expiration Date** 10/31/2014

**Tank Pressure** 512

**Barometric Pressure** 747 mmHg

**Blow Sample Number** 1      **Blow Duration** 3.39 sec

**Blow Volume** 1771 cc      **End-of-Blow Time** 10:26

**Blow Sample Number** 2      **Blow Duration** 3.41 sec

**Blow Volume** 1745 cc      **End-of-Blow Time** 10:29

**Tamper Evident Stamp** 06eaecd3

**Test Status Code** 0

**Test Status** Success

thus  
SD

**Intox EC/IR-II: Accuracy Check**

*Department of Forensic Science*

Serial Number: 010603      Test Number: 1377

Test Date: 08/26/2014      Test Time: 10:31 EDT

Dry Gas Target: 0.098

Lot Number: AG230501-37      Exp Date: 10/31/2014

Tank Pressure: 507 psi      Barometric Pressure: 747 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	10:32
CHK	0.098	10:33
BLK	0.000	10:35
CHK	0.098	10:35
BLK	0.000	10:37
CHK	0.098	10:37

Test Status: *Success*

Calibration CRC: CBA49532

LMS  
Jed

**Intox EC/IR-II: Accuracy Check**

*Department of Forensic Science*

Serial Number: 010603      Test Number: 1378

Test Date: 08/26/2014      Test Time: 10:49 EDT

Dry Gas Target: 0.295

Lot Number: AG421201-01      Exp Date: 07/31/2016

Tank Pressure: 350 psi      Barometric Pressure: 747 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	10:49
CHK	0.293	10:50
BLK	0.000	10:52
CHK	0.292	10:53
BLK	0.000	10:54
CHK	0.291	10:55

Test Status: *Success*

Calibration CRC: CBA49532